



## Short title: Anaesthesia machine safety BP

### 1. Purpose

Anaesthesia machines are fundamental to anaesthesia. They have continued to evolve in sophistication and functionality, becoming increasingly diverse and complex. ANZCA professional document *PS54(A) Position statement on the minimum safety requirements for anaesthesia machines and workstations for clinical practice* seeks to ensure that the minimum standards accord with advances in technology. The safety requirements are intended to facilitate decision-making regarding the need to upgrade, replace, or retain anaesthesia machines in current use. The requirements should be interpreted in the context of relevant Australian and New Zealand standards.

### 2. Scope

This professional document is intended to apply to all anaesthesia machines used in the hospital environment. This document is not intended to apply to anaesthesia machines used in the field, nor is it intended to apply to ventilators or monitoring systems, which are often integrated in modern machines.

### 3. Background

A review of *PS54(A)*, then known as *T03 Minimum safety requirements for anaesthetic machines for clinical practice*, commenced in 2010. A meeting at ANZCA House on October 8, 2010 of reviewers appointed by ANZCA Council, Fellows with an interest in draw-over and disaster relief anaesthesia apparatus, and others with specific standards interest and experience, informed the review. The Faculty of Pain Medicine Board, ANZCA regional and national committees, and the ANZCA Trainee Committee were invited to comment on the proposed document. In accordance with *CP24(G) Policy for the development and review of professional documents*, the resulting document was subject to a pilot phase, with feedback from stakeholders encouraged. In July 2011, during this pilot phase, the deadline for compliance was extended from January 2012 to January 2013 to allow manufacturers time to either develop the required modifications or additions to their machines, or to source adequate alternatives from other manufacturers which were compatible with their machines. Other amendments made at this time related to a high pressure relief valve or other means of automatically preventing dangerous high and/or prolonged pressures in the breathing system, and requirements for switching “off” an electronic anaesthesia machine.

As a result of the feedback received during the pilot phase, which required considerable deliberation, ANZCA Council agreed, in August 2012, to extend the timeline for compliance with the final revision. The extension of one year to January 1, 2014 was for the purpose of allowing manufacturers time to either develop the required modifications or additions to their machines, or to source adequate alternatives from other manufacturers, which are compatible with their machines. At this time the document was reissued as a professional standard, *PS54(A) Position statement on the minimum safety requirements for anaesthesia machines and workstations for clinical practice*, reflecting the decision to abolish the technical category of professional documents.

A further abbreviated review was undertaken during 2020 to address an apparent inconsistency between the accompanying guideline and an ISO standard referring to the requirement for the presence of a display of gas supply line pressures.

## 4. Issues

The document has been restructured for clarity: “mandatory” and “recommended” safety requirements have been amalgamated; “other” safety requirements are now detailed under “maintenance” requirements. Terminology has been aligned with the relevant standards documents. The issues of high and low pressures in the patient’s airway have been separated across two items (now items 3.16 and 3.17).

It was suggested that a statement, enforcing manufacturers’ requirements in relation to equipment testing, should be included in the document. The document development group, however, did not wish to restrict department or anaesthetist choice with respect to combinations of equipment and consequently this was not incorporated. In addition it was considered that the recommended testing of the anaesthesia machine by the anaesthetist before use would be sufficient to ensure the safe functioning of the machine.

At the close of the pilot phase, the title of the document was amended, consistent with the broader suite of ANZCA professional documents.

Item 3.4 was amended in the most recent review to ensure that PS54(A) aligns with the wording contained in ISO standard 201.101.4.3.

## 5. Summary

Anaesthesia machines are fundamental to anaesthesia and have continued to evolve in sophistication and functionality, as well as diversity and complexity. This professional document seeks to ensure that the minimum standards keep pace with advances in technology, and to assist in selecting, maintaining and replacing anaesthesia machines.

### Document development group

Professor Kate Leslie (chair), FANZCA, Councillor  
Professor Barry Baker, FANZCA, FCICM, Executive Director of Professional Affairs  
Dr Patrick Farrell, FANZCA, Councillor  
Dr Peter Roessler, FANZCA, Director of Professional Affairs (Professional Documents)

The document development group acknowledges the technical contribution of Mr Steven Threlfo.

### Consultation

The following individuals participated in the meeting held at ANZCA House on October 8, 2010:

Professor Kate Leslie (chair), FANZCA, President  
Professor Barry Baker, FANZCA, FCICM, Executive Director of Professional Affairs  
Mr John Biviano, Director, Policy, Quality and Accreditation  
Ms Rebecca Conning, Policy Officer  
Dr Mark Fajgman, FANZCA  
Dr Patrick Farrell, FANZCA, Councillor  
Dr Phoebe Mainland, FANZCA  
Dr George Merridew, FANZCA, FFPANZCA  
Dr Haydn Perndt, AM, FANZCA  
Dr Peter Roessler, FANZCA, Director of Professional Affairs (Professional Documents)

The following were also consulted:

Faculty of Pain Medicine Board  
ANZCA Trainee Committee

ANZCA regional/national committees  
Australian anaesthesia machine manufacturers and distributors

*Professional documents of the Australian and New Zealand College of Anaesthetists (ANZCA) are intended to apply wherever anaesthesia is administered and perioperative medicine practised within Australia and New Zealand. It is the responsibility of each practitioner to have express regard to the particular circumstances of each case, and the application of these ANZCA documents in each case. It is recognised that there may be exceptional situations (for example, some emergencies) in which the interests of patients override the requirement for compliance with some or all of these ANZCA documents. Each document is prepared in the context of the entire body of the college's professional documents, and should be interpreted in this way.*

*ANZCA professional documents are reviewed from time to time, and it is the responsibility of each practitioner to ensure that he or she has obtained the current version which is available from the college website ([www.anzca.edu.au](http://www.anzca.edu.au)). The professional documents have been prepared having regard to the information available at the time of their preparation, and practitioners should therefore take into account any information that may have been published or has become available subsequently.*

*Whilst ANZCA endeavours to ensure that its professional documents are as current as possible at the time of their preparation, it takes no responsibility for matters arising from changed circumstances or information or material which may have become available subsequently.*

Promulgated: 2013  
Current document: May 2021

© Copyright 2021 – Australian and New Zealand College of Anaesthetists. All rights reserved.

*This work is copyright. Apart from any use as permitted under the Copyright Act 1968, no part may be reproduced by any process without prior written permission from ANZCA. Requests and inquiries concerning reproduction and rights should be addressed to the Chief Executive Officer, Australian and New Zealand College of Anaesthetists, 630 St Kilda Road, Melbourne, Victoria 3004, Australia. Email: [ceo@anzca.edu.au](mailto:ceo@anzca.edu.au)*

ANZCA website: [www.anzca.edu.au](http://www.anzca.edu.au)